

## Correction: *Evaluation of intervertebral body implant performance using active surveillance of electronic health records*

Frankenberger EA, Resnic FS, Ssemaganda H, *et al.* Evaluation of intervertebral body implant performance using active surveillance of electronic health records. *BMJ Surg Interv Health Technologies* 2022;4:e000125. doi: 10.1136/bmjsit-2021-000125

The highlighted data in Table 3 was updated.

Table 3 IPTW analysis-estimated ORs for the risk of adverse outcomes with treatment using the CONCORDE Bullet					
Outcomes: IPTW analysis	OR	Lower CI	Upper CI	SE	P value
Primary endpoint:					
Reoperation at 1 year	1.01	0.99	1.03	0.01	0.27
Secondary endpoints:					
Mortality at 1 year	1.01	0.99	1.02	0.01	0.40
Any hospitalization at 1 year	1.07	0.98	1.16	0.04	0.11
Blood transfusion (postoperative)	0.92	0.83	1.01	0.05	0.08
Surgical site infection	n/a	n/a	n/a	n/a	n/a
IPTW, inverse probability of treatment weighting.					

**Open access** This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

© Author(s) (or their employer(s)) 2022. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

*BMJ Surg Interv Health Technologies* 2022;4:e000125corr2. doi:10.1136/bmjsit-2021-000125corr2

